Test Mixture

for GC

GC-192001-CS-E

January 2025

GROB Test Mixture P/N 192001

Solvent :		n-Hexane	
Substances:	1.	2,3-Butanediol	532 mg/L
	2.	n-Decane	284 mg/L
	3.	n-Undecane	285 mg/L
	4.	n-Octanol	354 mg/L
	5.	2-Ethylhexanoic acid	377 mg/L
	6.	Nonanal	400 mg/L
	7.	2,6-Dimethylphenol	330 mg/L
	8.	2,6-Dimethylaniline	330 mg/L
	9.	Decanoic acid methyl ester	430 mg/L
	10.	Undecanoic acid methyl ester	419 mg/L
	11.	Dicyclohexylamine	312 mg/L
	12.	Dodecanoic acid methyl ester	412 mg/L

This test mixture is not suitable for calibration.
It is intended for checking the elution behavior of GC columns.

We assume no liability for correctness of the concentration values.



CS - Chromatographie Service GmbH, Am Parir 27, 52379 Langerwehe/Germany
Phone: +49 2423 / 40493-0 · E-Mail: info@cs-chromatographie.de